



**delta 9-THC (parent), delta 9-HYDROXY THC, CARBOXY- delta 9-THC IN WHOLE BLOOD FOR GC/MS CONFIRMATIONS  
USING: 100 mg STYRE SCREEN<sup>®</sup> SSTHC  
EXTRACTION COLUMN**

Part #: SSTHC116

**1. PREPARE SAMPLE:**

To 1-2 mL whole blood add appropriate internal standards prepared in alcohol.

Add drop-wise 2 mL ice cold acetonitrile.\*

Mix thoroughly and centrifuge.

Decant acetonitrile into a clean tube. Evaporate acetonitrile under a stream of air or nitrogen to ~ 200uL.

Add 2 mL distilled water (pH~6.0-7.0)

(The sample is ready to be extracted.)

**\*NOTE:** The acetonitrile should be cold (recommended storage in freezer at <0 C just prior to use) and it should be added very slowly to ensure proper mixing of organic phase with the whole blood. If added too quickly, the blood could precipitate too fast possibly resulting in lower recoveries.

**2. APPLY SAMPLE:**

Load sample directly to column without any preconditioning.

**3. WASH COLUMN:**

Wash with 1 mL (84/15/1) Water/ Acetonitrile/ NH<sub>4</sub>OH.

Dry column thoroughly under vacuum (10 mm Hg) or positive pressure (~ 80-100 psi) for 10-15 minutes.

**NOTE:** (It is important to dry the column properly to achieve the highest recovery of all compounds. Any residual moisture will slow down the drying of the elution solvents prior to derivatization. Also, any residual moisture could reduce the reactivity of the derivatizing agent.)

**4. ELUTE THC, THC-OH, THC-COOH**

1 x 3 mL Hexane/ Ethyl Acetate/ Glacial Acetic Acid (49: 49:2)

Collect at 1-2 mL/ minute.

## 5. DRY ELUATE:

Evaporate fraction(s) to complete dryness under stream of dry air or nitrogen at <40°C

## 6. DERIVITIZE

Add 50 µL ethyl acetate, vortex mix

Add 50 µL BSTFA (with 1% TMCS).

or

Add 50 µL MTBSTFA (with 1% TBMCS).

Mix/vortex.

React 20 minutes at 70°C.

Remove from heat source to cool.

**NOTE:** Do not evaporate BSTFA.

## 7. QUANTITATE

Inject 2 µL onto gas chromatograph.

For MSD monitor the following ions:

## DERIVATIZATION PROCEDURE

<u>Derivatizing Agent</u>	THC {T-005} <sup>**</sup> <u>(D3 THC) {T-003}<sup>**</sup></u>	THC-OH {H-041} <sup>**</sup> <u>(D3 THC-OH) {H-027}<sup>**</sup></u>	THC-COOH {T-006} <sup>**</sup> <u>(D9 THC-COOH) {T-007}<sup>**</sup></u>
<b>BSTFA</b>	371, 343, 386 (374, 346, 389)	371, 459, 474 (374, 462, 477)	371, 473, 488 (380, 479, 497)
<b>MTBSTFA</b>	371, 428, 345 (374, 431, 348)	413, 369, 501 (416, 372, 504)	413, 515, 572 (422, 524, 581)

\* Suggested internal standard for GC/MS: D9-Carboxy-delta<sup>9</sup>-THC, D<sup>3</sup>-Hydroxy- delta<sup>9</sup>-THC, D<sup>3</sup>-delta<sup>9</sup>-THC

\*\* Ions common to deuterated delta-9 THC and non-deuterated compounds.